

User Manual

Smart Blood Pressure Monitor



IP21

MD

SKU TMB-2088

REF TMB-2288-B



SKU TMB-2088

REF TMB-2288-B

Guangdong Transtek Medical Electronics Co., Ltd.

Zone A, No.105, Dongli Road, Torch Development District, 528437 Zhongshan, Guangdong, China

CE 0123

EC REP MDSS GmbH
Schiffgraben 41, 30175 Hannover, Germany

20240517-v6 TMB-2288-B UK

- Thank you for selecting the Kinetik Wellbeing Smart Blood Pressure Monitor.
- Please read the user manual carefully and thoroughly, this is to ensure the safe usage of the product. Please keep the user manual for future reference.

Table of Contents

SUPPORT	2
INTRODUCTION	3
• General Description	
• Indications for Use	
• Measurement Principle	
• Receiving and Inspecting your Monitor	
• Safety Information	
• Display and Symbols	
• Name of Each Part	
• Contents/Product Includes	
BEFORE YOU START	10
• Choice of Power Supply	
• Installing and Replacing the Batteries	
• Setting the Date and Time	
• Setting a user ID	
• Pair a smart device with the monitor	
MEASUREMENT	17
• Applying the cuff	
• Start a measurement	
DATA MANAGEMENT	21
• Recall the Records	
• Delete the Records	
INFORMATION FOR USER	26
• Tips for measurement	
• Maintenances	
ABOUT BLOOD PRESSURE	28
• What are systolic pressure and diastolic pressure?	
• What is the standard blood pressure classification?	
• Irregular pulse rate detector	
• Why does my blood pressure fluctuate throughout the day?	
• Why do I get a different blood pressure at home compared to the hospital?	
• Is the result the same if measuring on the right arm?	
TROUBLESHOOTING	31
SPECIFICATIONS	32
AUTHORIZED COMPONENT	33
EMC GUIDANCE	34

♥ Support

Our manual should provide you with all the information you need to set up and use this product.

If you have a question, have a look at our Troubleshooting page!

For further assistance, why not contact our Customer Care team directly? We're here to help!

Our Customer Care team are available from 9am-5pm, Monday to Friday (excluding bank holidays).

We promise to respond to all queries and will ensure to resolve any issue you may be having.

You can reach us by...

Live Chat:

Simply visit www.kinetikwellbeing.com and send us a message.

Email:

customercare@kinetikwellbeing.com

♥ General Description

Thank you for selecting the Kinetik Wellbeing Smart Blood Pressure Monitor. The monitor features blood pressure measurement, pulse rate measurement and the result storage.

The blood pressure monitor provides you with two years of reliable service.

FEATURES:

- **124 mm × 76 mm Digital LCD display**
- **Systolic, Diastolic Blood Pressure & Pulse Display**
- **Date / Time Indication**
- **Irregular pulse rate detection**
- **2 × 250 memory storage, Guest mode available**
- **Up-to-date measuring-during-inflation technology**

♥ Indications for Use

The Kinetik Wellbeing SmartBlood Pressure Monitor is a digital monitor intended for use in measuring blood pressure and pulse rate with arm circumference ranging from 22cm to 32cm (about 8¾"-12½") , 22cm to 42cm (about 8¾"-16½") or 22cm to 45cm (about 8¾"-17¾").

It is intended for adult indoor use only.

Only the matched cuff of 22cm to 45cm (about 8¾"-17¾") is applicable to pregnant women (including pre-eclampsia patients) and diabetic patients.

♥ Measurement Principle


















This product uses the Oscillometric Measuring method to detect blood pressure. Before every measurement, the unit establishes a "zero pressure" equivalent to the atmospheric pressure. Then it starts inflating the arm cuff, meanwhile, the unit detects pressure oscillations generated by beat-to-beat pulsatile, which is used to determine the systolic and diastolic pressure, and also pulse rate.

♥ Receiving and Inspecting your Monitor

Check that the device packaging has not been tampered with and make sure that all contents are present. Before use, ensure that there is no visible damage to the device or accessories and that all packaging material has been removed. If you have any doubts, do not use the device and contact Kinetik Wellbeing.

♥ Safety Information

The symbols below might be in the user manual, labeling or other component. They are the requirement of standard and using.

	Manufacturer		Type BF applied part
	Date of manufacture and made in China		Direct Current
			Serial Number
	Recyclable		Medical Device
	For indoor use only		Class II Equipment
	Consult instructions for use or consult electronic instructions for use		Batch code
	Refer to instruction manual/booklet To signify that the instruction manual/ booklet must be read. Note: The background color of the symbol is blue.		
	Caution Indicates that caution is necessary when operating the device or control close to where the symbol is placed, or that the current situation needs operator awareness or operator action in order to avoid undesirable consequences.		
	The symbol indicates that the product should not be discarded as unsorted waste but must be sent to separate collection facilities for recovery and recycling.		
	CE marking indicates that a product has been assessed by the manufacturer and deemed to meet EU safety, health and environmental protection requirements. It is required for products manufactured anywhere in the world that are then marketed in the EU.		
	Authorized representative in the European Community/ European Union		
	Catalogue Number / Model Number		

Precaution

- * This device is intended for indoor, home use and is not intended for public use.
- * This device is portable, but it is not intended for use during patient transport.
- * This device is not suitable for continuous monitoring during medical emergencies or operations.
- * This device is intended for non-invasive measuring and monitoring of arterial blood pressure. It is not intended for use on extremities other than the arm, or for any purpose other than obtaining a blood pressure measurement.
- * This device is for adults. Do not use this device on neonates or infants. Do not use it on children and adolescents unless otherwise instructed by a healthcare professional.
- * Consult with a healthcare professional before using this monitor if you suffer from the following conditions: common arrhythmias such as premature ventricular beats or atrial fibrillation; peripheral arterial disease; implantation with electrical devices; undergoing intravascular therapy; arteriovenous shunt or mastectomy.
Please note that any of these conditions may affect measurement readings, in addition to patient motion, trembling or shivering.
- * Do not use this device for diagnosis or treatment of any health problem or disease. Contact a healthcare professional if you have or suspect any medical problem. Do not change your medications without the advice of a healthcare professional.
- * If you are taking medication, consult a healthcare professional to determine the proper time to measure your blood pressure.
- * This device may be used only for the intended use described in this manual, the manufacturer shall have no liability for any incidental, consequential, or special damages caused by misuse or abuse.
- * Please use the device under the environment which is provided in the user manual. Otherwise, the performance and lifetime of the device will be impacted and reduced.
- * The device may require up to 30 minutes to warm up / cool down from the minimum/ maximum storage temperature before it is ready for use.
- * The blood pressure monitor, its adapter, and the cuff are suitable for use within the patient environment.
- * Do not wash the cuff in a washing machine or dishwasher.
- * The device contains sensitive electronic components. To avoid measurement errors, avoid taking blood pressure measurements near a strong electromagnetic field radiated interference signal or electrical fast transient/burst signal.
- * Wireless communication equipment, such as wireless home network devices, mobile phones, cordless telephones and their base stations, walkie-talkies may cause interference that may affect the accuracy of measurements. A minimum distance of 1 foot (30 cm) should be kept from such devices during a measurement.
- * This blood pressure monitor is intended for use by medical or non-medical staff. A patient can also be the intended user or operator.

Caution

- * Do not attempt to repair the unit yourself if it malfunctions. Only have repairs carried out by authorised service centers.
- * It is recommended that the performance should be checked after repair, maintenance, and every two years of use, by retesting the requirements in limits of the error of the cuff pressure indication and air leakage (testing at least at 50 mmHg and 200 mmHg). Please contact Kinetik Wellbeing for authorised service personnel.
- * Store your device, cuff and adapter in a clean and dry place, protect it against extreme moisture, heat, lint, dust and direct sunlight. Never place any heavy objects on it.
- * Make sure the rubber tube of the cuff is not squeezed, stretched, or kinked during storage.
- * Dispose of accessories, detachable parts, and the device according to any local guidelines.


Warning

- * Do not apply the cuff on an arm that has an intravenous drip or a blood transfusion attached.
- * Do not kink, fold, stretch, compress, or otherwise deform the tube during measuring, as the cuff pressure might continuously increase, which could prevent blood flow and result injury.
- * Taking blood pressure measurements too frequently could disrupt blood circulation and cause injuries.
- * Do not apply cuff to areas on patient where skin is delicate or damaged. Check cuff site frequently for irritation.
- * Do not place the cuff on the arm of a person whose arteries or veins are undergoing medical treatment, i.e. intra-vascular access or intra-vascular therapy or an arteriovenous (A-V) shunt, which could disrupt blood circulation and cause injuries.
- * Do not place the cuff on the arm on the same side of a mastectomy (especially when lymph nodes have been removed), it is recommended to take measurements on the unaffected side.
- * Do not wrap the cuff on the same arm to which another monitoring device is applied. One or both devices could temporarily stop functioning if you try to use them at the same time.
- * Please check (for example, by observation of the limb concerned) that the operation of the device does not result in prolonged impairment of patient blood circulation.
- * On the rare occasion of a fault causing the cuff to remain fully inflated during measurement, loosen and remove the cuff immediately. Prolonged high pressure applied to the arm (cuff pressure >300 mmHg or constant pressure >15 mmHg for more than 3 minutes) might lead to bruising and discolored skin.
- * Do not use this device with high-frequency (HF) surgical equipment at the same time.
- * This device is not used in conjunction with oxygen rich environments, not intended for use with flammable anaesthetics, not intended for use in conjunction with flammable agents.
- * Do not touch output of the batteries/adaptor and the user simultaneously.
- * The power cord is considered the disconnect device for isolating this equipment from supply mains.
- * Do not position the equipment so that it is difficult to reach or disconnect.
- * Do not use this device if you are allergic to polyester, nylon, or plastic.
- * Only use accessories approved by manufacturer. Using unapproved accessories might cause damage to the unit and injure users.
- * If you experience discomfort during a measurement, such as pain in the arm or other complaints, press the Power button immediately to release the air from the cuff.
- * Do not use the device while under maintenance, or being serviced.
- * The air tube poses a risk of strangulation. Furthermore, the small parts of product and batteries present a choking hazard if swallowed. They should therefore always be kept away from infants/children.
- * Sensor degradation or looseness may reduce performance of device or cause other problems.

Notice

- * You can use this device to take your own measurement, no third-party operator is required.
- * Adapter is specified as a part of the ME (Medical Electrical) Equipment.
- * At the request of authorised service personnel, circuit diagrams, component part lists, descriptions, and calibration procedures will be made available by Kinetik Wellbeing.
- * The expected lifetime of the cuff may vary by the frequency of washing, skin condition, and storage state.
- * Please report to the manufacturer and the competent authority of the country in which you are established about any serious incident that has occurred in relation to this device.


♥ Display and Symbols



SYMBOL	EXPLANATION
1	Systolic blood pressure reading
2	Blood pressure level See the blood pressure classification on page 27-28.
3	Average value symbol Appears when viewing the average value of the latest 7 days under the memory mode.
4	Diastolic blood pressure reading
5	Excessive body motion detector symbol Appears when talking, moving, or shaking of the arm with the cuff on is detected during a measurement. NOTE: The measured blood pressure reading may not be accurate when this symbol is displayed with the reading.
6	Cuff wearing OK symbol Appears when the cuff is wrapped well.
7	Pulse rate symbol Flashes when detected during a measurement.
	Irregular pulse rate symbol Appears when detected during a measurement. Refer to page 28 for more information.

SYMBOL	EXPLANATION
8	Pulse display Pulse rate appears after the measurement.
9	Warning Messages Refer to page 30 for more information.
10	<p>Blood Pressure Trends Appears when viewing the blood pressure trends of the specified period of the past 7 days under the memory mode. The X-axis of the histogram represent the time, while the Y-axis represents the value of systolic and diastolic pressure. Refer to page 20 for more information.</p>
11	<p>Data pending to transmit symbol Displays when the data transmission failed. The measurements are saved on the device and will be sent when a successful connection is achieved.</p>
12	<p>Bluetooth symbol Flashes when the Bluetooth is working.</p>
13	<p>Adapter power mode / Adapter error symbol Appears when the power is supplied by AC adapter. Or appears along with "bAt H& " to indicate the voltage of the blood pressure monitor is high.</p>
14	<p>Low Battery symbol Indicate the battery is low when appears with "BAT Lo & ".</p>
15	Date / Time display
16	<p>User ID symbol The user ID number appears when the monitor is operated by the selected user.</p>
17	<p>Memory symbol Indicate it is in the memory mode and which the group of memory it is.</p>

♥ Name of Each Part




♥ Contents/Product Includes

- Blood Pressure Monitor (TMB-2288-B)
- Cuff (Type BF applied part)
Upper arm cuff: 22-45cm.
- User manual
- 4x AA batteries
- Type-C cable (not included)
- AC Adapter (not included)
Reference model: BLJ06L050100U-B (with UK plug)

♥ Choice of Power Supply

1. Battery powered mode:
6VDC 4×AA batteries
2. AC adaptor powered mode:
5V \equiv 1A
(Please only use AC adaptor authorized by manufacturer-AC adaptor is not included).
Please unplug the adaptor to stop using mains electricity.






⚠ CAUTION

Please use the correct batteries and approved power and approved power adaptor which complies with local safety standards.

♥ Installing and Replacing the Batteries

- Slide off the battery cover.
- Install or replace 4 x AA size batteries according to the polarity indications inside the battery compartment.
- Place back the battery cover.



Any time the battery is low, the LCD will display the icon , "bAt Lo" &  and then power off automatically after about 5 seconds.

Replace the batteries:

- When the low battery symbol appears. (see the LCD display on the right picture)
- When any button is pressed and nothing is displayed on the screen.




⚠ CAUTION

- New and used batteries, or different types of batteries shall not be used together.
- Remove batteries if the device is not likely to be used for some time.
- Do not heat or deform the batteries, or dispose of them in fire.
- Batteries should not be disposed of with household waste.
- Please check with your local authority for battery recycling advice.

♥ Setting the Date and Time

It is important to set the date and time before using your blood pressure monitor for the first time, so that a correct time stamp can be assigned to each record that is stored in the memory.

(The setting range of the year: 2022—2052, Time format: 24H/12H)

1. When the monitor is off, press and hold the "START/STOP" button, it will display the bluetooth symbol  first.
Press the "START/STOP" button again, it will enter the [Year] setting. Press the "USER" or "MEM" button to change the year.
Press and hold the "USER" button to quickly advance the years.
Press and hold the "MEM" button to quickly go backwards through the years.



2. Press the "START/STOP" button to confirm the [Year], then the date format will flash. Press the "USER" or "MEM" switch the date format between [month/day] and [day/month].



3. Press the “START/STOP” button to confirm the date format, then the [Month] will flash. Repeat the same steps to set the [Month] and [Day].



4. Repeat the same steps to set the time format between [24H] and [12H].



5. Repeat the same steps to set the [Hour] and [Minute].










6. After confirming the [Minute], the LCD will display all the settings you have done, and then it will display “do nE” and the device will turn off automatically.

Note:

1. The date and time will automatically be set when the monitor is paired with app. Refer to the section “Pair a smart device with the monitor”.
2. The device can also enter the setup mode automatically when the battery is installed for the first time.

♥ Setting a user ID

There are 3 user ID    available. The user  and , have 250 memories, so that 2 people can save their measured values separately. The user , has no memory space, is reserved for guest.




1. When the monitor is off, press the “USER” button to display the current user ID. Press the “USER” button again to switch the user ID among the user ,  and .





2. Press the “START/STOP” button to confirm the selected user ID, it will enter the measurement automatically. Press the “MEM” button to confirm the selected user ID, it will enter the memory query automatically.

♥ Pair a smart device with the monitor

You can measure your blood pressure and have the measurement data automatically sent to a smart device (such as smartphone or tablet) with Bluetooth wireless connectivity.

1. Download the Blood Pressure Diary by MedM app  or the MedM Health app  from the App store or Google Play.
2. Enable Bluetooth and the MedM Health app on your smart device. Make sure both Bluetooth and app are ON when pair-up is proceeding.
3. When the monitor is off, press and hold the “MEM” button to start pair-up, the bluetooth symbol  will flash.



4. If successful, the bluetooth symbol  will not flash any more and the monitor will automatically shut off after several seconds.
5. In the app, select the menu in the top left corner of the screen. Then select 'My Devices'.
6. At the bottom of the screen select 'Add New'. This will automatically start scanning for a new device. When detected, select the TMB-2288-B.
7. Once selected you will be asked if you would like to receive data automatically. If no, you will need to manually input results. If yes, you will then be asked if you want to receive data in the background.

Note:

1. The date and time on your monitor will automatically be set after paired with your smart device successfully.
2. The device can also enter the Bluetooth pairing automatically when the battery is installed for the first time.

If unsuccessful within 60 seconds, it is judged timeout and the monitor will shut off.

Specifications for Bluetooth Transmission		
Bluetooth	Throughput	2.5K-5K
	Latency	50ms
	PER	< 10%
	Operating Frequency	2400-2480MHz
	Transmission Power	0dBm
	Transmission Distance	10m

Note:

- The necessary Quality of Service (QoS) is fully considered here for wirelessly enabled functions.
- Interference may occur in the vicinity of equipment marked with the following symbol (Ⓜ). And TMB-2288-B may interfere the vicinal electrical equipment.
- Keep the monitor at least 20 centimeters away from the human body (especially the head) when data transmission is proceeding after measurement.
- To enable the data transmission function, this device shall be paired to an appropriate BT mobile terminal.

Warning

About a wireless communication interference

The monitor operates in the unlicensed ISM band at 2.4 GHz. In case it is used around the other wireless devices including microwave and wireless LAN, which operate at the same frequency band as the monitor, there is a possibility that interference occurs between the monitor and such other devices. If such interference occurs, please stop the operation of other devices or relocate the monitor before using it or do not use it around the other wireless devices.

List of compatible devices:

For iOS devices:

The operating system must be iOS 13.0 or more.

For Android devices:

The operating system must be Android 5.0 or more.

♥ Applying the cuff

Only use a cuff that has been approved by Kinetik Wellbeing for this device. Before use, please confirm if it fits your arm circumference.

1. Remove all jewelry , such as watches and bracelets from your left arm.

Note: If your doctor has diagnosed you with poor circulation in your left arm, use your right arm.

2. Roll or push up your sleeve to expose the skin. Make sure your sleeve is not too tight.

3. Hold your arm with your palm facing up and tie the cuff on your upper arm, align the Artery indicator Φ with the main Artery (on the inside of your arm).

Note: Locate the main Artery by pressing with 2 fingers approximately 2 cm above the bend of your elbow on the inside of your left arm. Identify where the pulse can be felt the strongest, that is your main Artery.

4. Make sure the bottom edge of the arm cuff 2 to 3 cm above the inside elbow. Then wrap the cuff securely.

Note: The cuff should be snug but not too tight. You should be able to insert one finger between the cuff and your arm.


5. Sit upright in a comfortable chair with your back against the backrest of the chair. Keep your feet flat and your legs uncrossed.

Place your arm resting comfortably on a flat table. The cuff worn on your arm should be placed at the same level as your right atrium of the heart.

6. Take 5-6 deep breaths and let's start measuring!

Helpful tips:

- Take the measurement in a silent room.
- Rest for 5 minutes before a measurement.
- Wait at least 3 minutes before another measurement. This allows your blood circulation to recover.
- Be relaxed and do not move and talk during the measurement procedure.
- For a meaningful comparison, try to measure under similar conditions. For example, take daily measurements at approximately the same time, on the same arm, or as directed by a healthcare professional.



♥ Start a measurement

You can use your monitor without pairing to a smart device. To pair your monitor with a smart device refer to page 15.


1. When the monitor is off, press the “START/STOP” button to turn on the monitor, and it will finish the whole measurement. Remain still and do not talk until the full measurement is complete.

(Take User 1 for example)


Cuff wrap detection




Inflating and measuring
(a pulse rate is detected)




Display and save
the measurement result



Note



- Any time, to stop the measurement, press the “START/STOP” button.
 - If you don't pair the device with the MedM Blood Pressure Diary or MedM Health app or don't keep the app ON, the bluetooth symbol  will still flash during the measurement.
2. Press the “START/STOP” button to turn off the monitor, or it will shut off within 1 minute.
 3. If your monitor is already paired with your smart device and both Bluetooth and app are ON, when the measurement has completed, the measurement result will start transmitting (only applicable to User 1 and User 2).

If successful, both symbols  and  will disappear and the monitor will automatically shut off.






If unsuccessful within 60 seconds, the monitor will shut off. In the case of a data transmission failure, up to 250 measurements are saved on the device and will be sent when a successful connection is achieved.

Note

- Both user  and  can store maximum 250 groups of record. When you pass that limit, every time you take a measurement, the monitor will prompt “FULL” first and the oldest record will be deleted from the list after the measurement.

4. About the irregular pulse rate and excessive body motion during the measurement.



- During a measurement, If an irregular pulse rate is detected, the symbol  will display in the measurement result. See page 28 for more information.
- During a measurement, when there is excessive body motion, especially on the arm the cuff is worn is detected, the symbol  will flash about 5 seconds and detect again. If excessive body motion is no longer detected, the symbol will disappear; if still detected, the symbol  will appear in the measurement result.


Note

The measured blood pressure reading may not be accurate if these symbol are displayed.



♥ Recall the Records

1. When the monitor is off, press the “MEM” button, the LCD will display the total groups of the memory records, with the user ID blinking.
2. Press the “USER” or “MEM” button to switch the user ID between user  and . Press the “START/STOP” button to confirm the selected user ID.
3. Then the LCD will display the average value of all the records of last 7 days and BP Trends chart of a specified period of last 7 days. (Example shown below for User 1)




Note

- BP trends is a histogram, designed to display the average blood pressure value of a specified period of last 7 days, to help the user to visually know their blood pressure changes.
- If there is no measurements for that period of time on a given day, the histogram for that day will be blank.
- For the BP trends chart, the X-axis of the histogram represent the time (from left to right, from Sunday, Monday, Tuesday, Wednesday, Thursday, Friday to Saturday), while the Y-axis of the histogram represents the average value of systolic (high) pressure and diastolic (low) pressure of the specified period each day.
- The data displayed on the top left is the maximum average systolic pressure value in the past 7 days, while the bottom left is the minimum diastolic pressure value in the past 7 days.

- When the maximum average systolic pressure is higher than 170mmHg or the maximum average diastolic pressure is higher than 110mmHg, the symbol“▲” will be displayed.
- When the minimum average systolic pressure is lower than or equal to 100mmHg, the minimum average diastolic pressure is lower than or equal to 70mmHg, the grid corresponding to 100mmHg or 70mmHg will be lit.

4. Press the “USER” or “MEM” button to display the next record.



The Time and Date will display alternately.

Note:


- If there is untransmitted data, the symbol displays on the record.
- The latest record (1) is shown first. Each new measurement is assigned to the first (1) record. All other records are pushed back one digit (e.g., 2 becomes 3, and so on), and the last record (250) will be dropped from the list.

♥ Delete the Records

Delete the records by following the steps below.

A: Delete one record (Example shown below for User 1)

1. Enter the memory recall mode as described in the previous section "Recall the records", find out the record you want to delete.
2. Press and hold the “MEM” button about 3 seconds, the LCD will display “dEL y” and blink.




3. Press the “START/STOP” button to confirm the deletion, the LCD will display “dEL donE” and then the previous record will be displayed.




Note

Before you confirm the deletion, you could press the “USER” or “MEM” button to switch the LCD display from “dEL y” to “dEL no”. Then press the “START/STOP” button, you could stop clearing the memory.




B: Delete all records (Example shown below for User 1)

1. Enter the memory recall mode as described in the previous section "Recall the records".
2. Press and hold the "USER" and "MEM" button for about 5 seconds, the LCD will display "dEL AL" and blink.



3. Press the "START/STOP" button to confirm the deletion, the LCD will display "dEL donE".




Note

Before you confirm the deletion, you could press the "USER" and "MEM" button to switch the LCD display from "dEL AL" to "dEL no". Then press the "START/STOP" button, you could stop clearing the memory.



4. Once deleted, your readings cannot be restored. The LCD will display "----" like the following picture.

Press the "START/STOP" button to turn off the monitor, otherwise it will power off automatically within 1 minute.



♥ Tips for Measurement

Measurements may be inaccurate if taken in the following circumstances.



Within 1 hour
after dinner or drinking



Immediate measurement
after tea, coffee, smoking



Within 20 minutes after taking a bath



When talking or moving your body



In a very cold environment



When you want to discharge urine



♥ Maintenance

In order to get the best performance, please follow the instructions below.

1. Cleaning Process:

- Step 1: Make sure to switch off and unplug the device prior to cleaning.
 Step 2: Use a soft cloth and dampen it with soapy water to clean the cuff first, and then use a soft cloth and dampen it with clear water to remove residual soap until there is no visible residual contaminants. Attention shall be paid to avoid liquid leaking into the cuff.
 Step 3: Use a dry soft cloth to wipe the cuff, in order to remove residual moisture.
 Step 4: Dry the cuff in a well-ventilated area after cleaning.

2. Disinfection Process:



- Step 1: Make sure to switch off and unplug the device prior to disinfection.
 Step 2: Use a soft cloth with 70% isopropanol to disinfect the cuff for about 10 minutes. Attention shall be paid to avoid liquid leaking into the cuff.
 Step 3: Use a clean dry cloth or towel to wipe off the disinfectant until there is no visible residue.
 Step 4: Dry the cuff in a well-ventilated area after disinfection.

• Suggestion:

Frequency of Cleaning and Disinfection:
 For single patient multiple use, it's recommended to clean the device surface once a month or whenever it's necessary.
 For multiple patient multiple use, it's recommended to clean the device every time before and after usage. Maintenance procedures shall be taken as per instruction.


♥ What are systolic pressure and diastolic pressure?

When ventricles contract and pump blood out of the heart, the blood pressure reaches its maximum value in the cycle, which is called systolic pressure. When the ventricles relax, the blood pressure reaches its minimum value in the cycle, which is called diastolic pressure.



♥ What is the standard blood pressure classification?

The blood pressure classification published by World Health Organization (WHO) and International Society of Hypertension (ISH) in 1999 is as follows:



Blood Pressure (mmHg) \ Level	Optimal	Normal	High-normal	Mild	Moderate	Severe
SYS	<120	120-129	130-139	140-159	160-179	≥180
DIA	<80	80-84	85-89	90-99	100-109	≥110


⚠ CAUTION

Only a healthcare professional can tell your normal BP range. Please contact a healthcare professional if your measuring result falls out of the range. Please note that only a healthcare professional can tell whether your blood pressure value has reached a dangerous point.

♥ Irregular Pulse Rate Detector

An irregular pulse rate will be detected if there is an irregular pulse rhythm while measuring systolic and diastolic blood pressure. When measurements were performed, the monitor will record all pulse intervals and calculate the average. If two or more pulse intervals were recorded, and the difference between each interval and the average is larger than $\pm 25\%$ of the average; or if four or more pulse intervals were recorded, and the difference between each interval and the average is larger than $\pm 15\%$ of the average value, the irregular pulse symbol will be displayed along with measurement results.

⚠ CAUTION

The appearance of the  icon indicates that a pulse irregularity consistent with an irregular pulse rate was detected during measurement. Usually this is NOT a cause for concern. However, if the symbol appears often, we recommend you seek medical advice. Please note that the irregular pulse rate detector only serves as a non-medical feature, and the results cannot be used directly for clinical judgement. Please seek medical advice from professionals before making any medical decisions.

♥ Why does my blood pressure fluctuate throughout the day?

1. Individual blood pressure varies multiple times everyday. It is also affected by the way you tie your cuff and your measurement position, so please take the measurement under the same conditions.
- 2.If the user takes medicine, the pressure will vary more.
- 3.Wait at least 3 minutes for another measurement.



What you need to pay attention to when you measure your blood pressure at home:

- If the cuff is tied properly.
- If the cuff is too tight or too loose.
- If the cuff is tied on the upper arm.
- If you feel anxious.
- Taking 2-3 deep breaths before beginning will be better for measuring.
- Advice: Relax for 4-5 minutes until you calm down.



♥ Is the result the same if measuring on the right arm?

It is ok for both arms, but there will be some different results for different people. We suggest you measure the same arm every time.

If any abnormality arises during use, please check the following points:


PROBLEM	SYMPTOM	CHECK THIS	REMEDY
No power	Display can not light up.	Batteries are depleted.	Replace with new batteries.
		Batteries are inserted incorrectly.	Insert the batteries correctly.
		Adapter is inserted incorrectly.	Insert the AC adapter correctly.
High Battery	Replace batteries & bAt H shows	The battery is too high.	Replace with new batteries.
Low Battery	Replace batteries & bAt Lo shows	The battery is too low.	Replace with new batteries.
Error message	Measure again & E 01 shows	The cuff is loose.	Re-wear the cuff before measurement.
	Keep quiet & E 02 or E 03 shows	Excessive body motion (such as shaking of the arm with the cuff on or weak Pulse is detected.)	Do not move during measurement and measure again.
	Measure again & E 03 shows	No pulse is detected.	Wear the cuff again before measurement.
	Device failure & E 04 shows	The measurement failed.	Relax for 5 minutes and measure again.
	EExx shows	Hardware error (XX can be some digital symbol, such as 01, 02, 03, etc.)	Turn off monitor and measure again. If EExx still appears on the display, please contact the retailer or our customer service.
	Replace batteries & bAt H shows	The voltage of the blood pressure monitor is high.	Replace with the authorized adapter.
Warning message	out shows	Out of measurement range.	Relax for a moment and then measure again. If the problem persists, contact your physician.

NOTE: If the product still does not work, contact Kinetik Wellbeing. Under no circumstance should you disassemble or attempt to repair the unit by yourself.

External dimensions	Approx.178 mm × 101 mm × 42.89 mm
Display mode	Digital LCD V.A.124 mm × 76 mm
Weight	Approx.343 g (Excluding the batteries and cuff)
Measurement mode	Oscillographic testing mode
Mode of operation	Continuous operation
Measurement range	Rated cuff pressure: 0 mmHg~299 mmHg Measurement pressure: SYS: 60 mmHg ~ 230 mmHg DIA: 40 mmHg ~ 130 mmHg Pulse value: (40-199) beat/minute
Accuracy	Static Pressure: 5°C-40°C within ±3mmHg Pulse value: ±5% Clinical validation: Mean difference within ±5mmHg; Standard deviation ≤8mmHg
Normal working condition	A temperature range of: +5°C to +40°C A relative humidity range of 15% to 90%, non-condensing, but not requiring a water vapour partial pressure greater than 50 hPa An atmospheric pressure range of 700 hPa to 1060 hPa
Storage & transportation condition	Temperature:-20°C to +60°C A relative humidity range of ≤ 93%, non-condensing, at a water vapour pressure up to 50 hPa An atmospheric pressure range of 500 hPa to 1060 hPa
Measurement perimeter of the upper arm	About 22-45cm Cuff model AC2245-05 is intended for adult population or those who are in pregnancy or suffer from diabetes.
Degree of protection	Type BF applied part
Protection against ingress of water	IP21 It means the device could be protected against solid foreign objects of 12,5mm Φ and greater, and against vertically falling water drops.
Device Classification	Battery Powered Mode: Internally Powered ME Equipment AC Adapter Powered Mode: Class II ME Equipment
Expected Lifetime	Device: 3 years or 30,000 measurements (may vary based on usage conditions) Cuff: 10000 times Alkaline battery: About 200-300 times
Types of use/reuse	Multiple patient multiple use

♥ Authorized Component

Please use the authorised adapter (not included).



Adapter

Type: BLJ06L050100U-B

Input: 100-240V, 50-60Hz, 0.2A max

Output: 5V --- 1000 mA

♥ Contact Information

For more information about our products, please visit
www.kinetikwellbeing.com

WARNING: No modification of this equipment is allowed.

NOTE: This blood pressure monitor complies with ISO 81060-2: 2018
(including pregnant women and diabetics).

♥ EMC Guidance

The ME EQUIPMENT or ME SYSTEM is suitable for home healthcare environments.

Essential performance:

Accuracy of measuring blood pressure and pulse rate

Measurement Range	Systolic pressure: 60-230 mmHg Diastolic pressure: 40-130 mmHg Pulse: 40-199 beats/minute
Rated Cuff Pressure	0-299 mmHg (0-39.9 kPa)
Accuracy	Pressure: ± 3 mmHg / 0.4 kPa Pulse: $\pm 5\%$

The Basis Safety of the Blood Pressure Monitor (TMB-2288-B) is as following:
Deviation from normal operation that poses an unacceptable risk to the patient or operator.

Warning: Don't be near the active HF surgical equipment and the RF shielded room of an ME system for magnetic resonance imaging, where the intensity of EM disturbances is high.

Warning: Use of this equipment adjacent to or stacked with other equipment should be avoided because it could result in improper operation. If such use is necessary, this equipment and the other equipment should be observed to verify that they are operating normally.

Warning: Use of accessories, transducers and cables other than those specified or provided by the manufacturer of this equipment could result in increased electromagnetic emissions or decreased electromagnetic immunity of this equipment and result in improper operation.

Warning: Portable RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30 cm (12 inches) to any part of the equipment including cables specified by the manufacturer. Otherwise, degradation of the performance of this equipment could result.

Technical description:

1. All necessary instructions for maintaining BASIC SAFETY and ESSENTIAL PERFORMANCE with regard to electromagnetic disturbances for the expected lifetime.
2. Guidance and manufacturer's declaration-electromagnetic emissions and Immunity.

Table 1

Guidance and manufacturer's declaration - electromagnetic emissions	
Emissions test	Compliance
RF emissions CISPR 11	Group 1
RF emissions CISPR 11	Class [B]
Harmonic emissions IEC 61000-3-2	Class A
Voltage fluctuations / flicker emissions IEC 61000-3-3	Comply

Table 2

Guidance and manufacturer's declaration – electromagnetic Immunity		
Immunity Test	IEC 60601-1-2 Test level	Compliance level
Electrostatic discharge (ESD) IEC 61000-4-2	±8 kV contact ±2 kV, ±4 kV, ±8 kV, ±15 kV air	±8 kV contact ±2 kV, ±4 kV, ±8 kV, ±15 kV air
Electrical fast transient/burst IEC 61000-4-4	±2 kV for power supply lines ±1 kV signal input/output 100 kHz repetition frequency	±2 kV for power supply lines Not Applicable 100 kHz repetition frequency
Surge IEC61000-4-5	±0.5 kV, ±1 kV differential mode ±0.5 kV, ±1 kV, ±2 kV common mode	±0.5 kV, ±1 kV differential mode Not Applicable
Voltage dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11	0% Ur; 0,5 cycle. At 0°, 45°, 90°, 135°, 180°, 225°, 270° and 315°. 0% Ur; 1 cycle and 70% Ur; 25/30 cycles; Single phase: at 0°. 0% Ur; 250 / 300 cycle	0% Ur; 0,5 cycle. At 0°, 45°, 90°, 135°, 180°, 225°, 270° and 315°. 0% Ur; 1 cycle and 70% Ur; 25/30 cycles; Single phase: at 0°. 0% Ur; 250 / 300 cycle
Power frequency magnetic field IEC 61000-4-8	30 A/m 50 Hz / 60 Hz	30 A/m 50 Hz / 60 Hz
Conducted RF IEC61000-4-6	3 V 0,15 MHz – 80 MHz 6 V in ISM and amateur radio bands between 0,15 MHz and 80 MHz 80% AM at 1 kHz	3 V 0,15 MHz – 80 MHz 6 V in ISM and amateur radio bands between 0,15 MHz and 80 MHz 80% AM at 1 kHz
Radiated RF IEC61000-4-3	10 V/m 80 MHz – 2,7 GHz 80% AM at 1 kHz	10 V/m 80 MHz – 2,7 GHz 80% AM at 1 kHz
NOTE U _T is the a.c. mains voltage prior to application of the test level.		

Table 3

Guidance and manufacturer's declaration - electromagnetic Immunity								
Radiated RF IEC61000-4-3 (Test specifications for ENCLOSURE PORT IMMUNITY to RF wireless communications equipment)	Test Frequency (MHz)	Band (MHz)	Service	Modulation	Maximum Power (W)	Distance (m)	IEC 60601-1-2 Test Level (V/m)	Compliance level (V/m)
	385	380-390	TETRA 400	Pulse modulation 18 Hz	1.8	0.3	27	27
	450	430-470	GMRS 460, FRS 460	FM ± 5k Hz deviation 1 kHz sine	2	0.3	28	28
	710	704-787	LTE Band 13, 17	Pulse modulation 217 Hz	0.2	0.3	9	9
	745							
	780							
	810	800-960	GSM 800/900, TETRA 800, IDEN 820, CDMA 850, LTE Band 5	Pulse modulation 18 Hz	2	0.3	28	28
	870							
	930							
	1720	1700-1990	GSM 1800; CDMA 1900; GSM 1900; DECT; LTE Band 1, 3, 4, 25; UMTS	Pulse modulation 217 Hz	2	0.3	28	28
	1845							
	1970							
	2450	2400-2570	Bluetooth, WLAN, 802.11 b/g/n, RFID 2450, LTE Band 7	Pulse modulation 217 Hz	2	0.3	28	28
	5240	5100-5800	WLAN 802.11 a/n	Pulse modulation 217 Hz	0.2	0.3	9	9
5500								
5785								